




NEUROCIRUGÍA

www.elsevier.es/neurocirugia


Clinical Research

Bilateral deep brain stimulation of the subthalamic nucleus: Targeting differences between the first and second side



Filipa de Oliveira^{a,b,c}, Rui Vaz^{a,b,c}, Clara Chamadoira^{a,b,c}, Maria José Rosas^{b,c,d}, Manuel J. Ferreira-Pinto^{a,c,e,*}

^a Department of Neurosurgery, Centro Hospitalar e Universitário de São João, Porto, Portugal

^b Department of Clinical Neurosciences, Faculty of Medicine, University of Porto, Portugal

^c Movement Disorders and Functional Neurosurgery Unit, Centro Hospitalar e Universitário de São João, Porto, Portugal

^d Department of Neurology, Centro Hospitalar e Universitário de São João, Porto, Portugal

^e Department of Surgery and Physiology, Faculty of Medicine, University of Porto, Portugal

ARTICLE INFO

Article history:

Received 21 March 2022

Accepted 18 July 2022

Available online 13 September 2022

Keywords:

Deep brain stimulation

Parkinson's disease

Subthalamic nucleus

Microelectrode recording

Stereotactic surgery

Functional neurosurgery

ABSTRACT

Introduction and objectives: Deep brain stimulation (DBS) of the subthalamic nucleus (STN) is a recognized treatment for drug-refractory Parkinson's disease (PD). However, the therapeutic success depends on the accuracy of targeting. This study aimed to evaluate potential accuracy differences in the placement of the first and second electrodes implanted, by comparing chosen electrode trajectories, STN activity detected during microelectrode recording (MER), and the mismatch between the initially planned and final electrode positions on each side.

Materials and methods: In this retrospective cohort study, we analyzed data from 30 patients who underwent one-stage bilateral DBS. For most patients, three arrays of microelectrodes were used to determine the physiological location of the STN. Final target location depended also on the results of intraoperative stimulation. The choice of central versus non-central channels was compared. The Euclidean vector deviation was calculated using the initially planned coordinates and the final position of the tip of the electrode according to a CT scan taken at least a month after the surgery.

Results: The central channel was chosen in 70% of cases on the first side and 40% of cases on the second side. The mean length of high-quality STN activity recorded in the central channel was longer on the first side than the second (3.07 ± 1.85 mm vs. 2.75 ± 1.94 mm), while in the anterior channel there were better MER recordings on the second side (1.59 ± 2.07 mm).

* Corresponding author.

E-mail address: manuel.jnfpinto@gmail.com (M.J. Ferreira-Pinto).

<https://doi.org/10.1016/j.neucir.2022.07.001>

1130-1473/© 2022 The Authors. Published by Elsevier España, S.L.U. on behalf of Sociedad Española de Neurocirugía. This is an open access article under the CC BY-NC-ND license (<http://creativecommons.org/licenses/by-nc-nd/4.0/>).

on the first side vs. 2.78 ± 2.14 mm on the second). Regarding the mismatch between planned versus final electrode position, electrodes on the first side were placed on average 0.178 ± 0.917 mm lateral, 0.126 ± 1.10 mm posterior and 1.48 ± 1.64 mm inferior to the planned target, while the electrodes placed on the second side were 0.251 ± 1.08 mm medial, 0.355 ± 1.29 mm anterior and 2.26 ± 1.47 mm inferior to the planned target.

Conclusion: There was a tendency for the anterior trajectory to be chosen more frequently than the central on the second side. There was also a statistically significant deviation of the second electrodes in the anterior and inferior directions, when compared to the electrodes on the first side, suggesting that another cause other than brain shift may be responsible. We should therefore factor this during planning for the second implanted side. It might be useful to plan the second side more anteriorly, possibly reducing the number of MER trajectories tested and the duration of surgery.

© 2022 The Authors. Published by Elsevier España, S.L.U. on behalf of Sociedad Española de Neurocirugía. This is an open access article under the CC BY-NC-ND license (<http://creativecommons.org/licenses/by-nc-nd/4.0/>).

Estimulación cerebral profunda bilateral del núcleo subtalámico: diferencias de posicionamiento entre el primer y el segundo lado

R E S U M E N

Palabras clave:

Estimulación cerebral profunda
Enfermedad de Parkinson
Núcleo subtalámico
Microregistro electrofisiológico
Cirugía estereotáxica
Neurocirugía funcional

Introducción y objetivos: La estimulación cerebral profunda (ECP) del núcleo subtalámico (NST) es reconocida como un tratamiento para la enfermedad de Parkinson (EP) refractaria al tratamiento farmacológico. Sin embargo, el éxito de esta intervención depende de la precisión de la colocación de los electrodos. Este estudio tuvo como objetivo evaluar las posibles diferencias de precisión entre la colocación del primer y segundo electrodo, comparando las trayectorias elegidas para cada lado, la actividad del NST detectada durante el microrregistro (MER) y la discrepancia entre las posiciones inicialmente planeadas y las finales.

Materiales y métodos: En este estudio retrospectivo analizamos datos de 30 pacientes sometidos a ECP bilateral. En la mayoría de los casos se usaron tres conjuntos de microelectrodos para determinar la ubicación fisiológica del NST. El posicionamiento final del electrodo estuvo asimismo condicionado por los resultados de la estimulación intraoperatoria. Se comparó la elección de canales centrales vs. no centrales. El vector euclidiano del desvío se calculó a partir de las coordenadas planeadas inicialmente y la posición final de la punta del electrodo, según una tomografía computarizada realizada al menos un mes después de la cirugía.

Resultados: La trayectoria central se eligió en 70% de los casos en el primer lado y en el 40% de los casos en el segundo lado. La duración media de la actividad de alta calidad del NST registrada en el canal central fue mayor en el primer lado que en el segundo ($3,07 \pm 1,85$ mm vs. $2,75 \pm 1,94$ mm), mientras que en el canal anterior hubo mejores registros de MER en el segundo lado ($1,59 \pm 2,07$ mm en el primer lado vs. $2,78 \pm 2,14$ mm en el segundo). En cuanto a la discrepancia entre la diana planeada y la posición final de los electrodos, los electrodos del primer lado se colocaron, por término medio, $0,178 \pm 0,917$ mm laterales, $0,126 \pm 1,10$ mm posteriores y $1,48 \pm 1,64$ mm inferiores al objetivo, mientras que los electrodos colocados en el segundo lado estaban $0,251 \pm 1,08$ mm mediales, $0,355 \pm 1,29$ mm anteriores y $2,26 \pm 1,47$ mm inferiores al objetivo planificado.

Conclusión: En el segundo lado, observamos una tendencia a elegir la trayectoria anterior más frecuentemente que la central. También hubo un desvío estadísticamente significativo de los segundos electrodos, en dirección anterior e inferior, en comparación con los electrodos del primer lado, lo que sugiere que la causa puede ser otra que no el *brain shift*. Por lo tanto, debemos tener esto en cuenta al planear la inserción del electrodo en el segundo lado. Podría ser útil planear el segundo lado más anteriormente, posiblemente reduciendo el número de trayectorias probadas por MER y la duración de la cirugía.

© 2022 Los Autores. Publicado por Elsevier España, S.L.U. en nombre de Sociedad Española de Neurocirugía. Este es un artículo Open Access bajo la licencia CC BY-NC-ND (<http://creativecommons.org/licenses/by-nc-nd/4.0/>).

Introduction

Deep brain stimulation (DBS) is an established treatment for advanced, drug-refractory Parkinson's disease (PD), comprising the application of chronic high-frequency electrical stimulation through an electrode implanted in a specific brain target.^{1,2} It produces a superior improvement in quality of life and reduction of motor symptoms' severity than isolated pharmacological therapy.^{3,4} A commonly chosen target in PD is the subthalamic nucleus (STN).

The therapeutic success of DBS depends on the accuracy of STN targeting.⁵⁻⁷ Literature shows that it is essential to target the dorsolateral, motor subdivision of the STN, avoiding penetration of its limbic and associative subdivisions (which might lead to executive function deficits and mood impairment), or the laterally located internal capsule (which could cause tonic muscular contractions and slurred speech).^{1,8} Thus, accurate STN targeting requires rigorous care at different steps, including planning and surgical execution.

However, stereotactic neurosurgery relies on the assumption that brain structures are static while events such as brain shift can hinder the correct placement of the electrodes, causing a discrepancy between the images acquired preoperatively and the actual location of structures during and after surgery.^{8,9} This movement of the brain is due to various factors, particularly the force of gravity, and the invasion of the intracranial space by air and the loss of cerebrospinal fluid that occur during the opening of the dura mater, causing an anterior pneumocephalus.¹⁰⁻¹² This explains why the shift generally occurs in a posterior direction when the patient is placed in a supine position during surgery.^{13,14} While some studies argue that shift of subcortical structures is very limited and does not significantly affect clinical outcome,^{10,15} Hunsche et al. defend that in bilateral DBS the brain shift that occurs after placing the first electrode should be considered relevant as it can affect the placement of the second one.¹⁶ Other studies have also shown that more adjustment with MER and intraoperative macrostimulation is usually required for the correct placement of the second electrode.^{6,17} It is therefore possible to have a more significant shift (thus increasing the distance between the planned target and the final electrode location) on the second implanted side, when compared to the first.⁷ The reduced accuracy in placement of the second electrode is related to a reduction in the threshold for adverse effects, and a less pronounced improvement in symptoms.^{18,19}

Taking this into account, this study aims to evaluate potential differences in targeting accuracy (defined as final electrode position equal to initially planned target), comparing the first and second implanted sides in patients undergoing bilateral STN-DBS surgery for PD. We hypothesize that if there are no differences in targeting accuracy between the first and second sides, then the frequency of choosing the central channel for final implantation will be similar for both. However, in case there are in fact differences, then this should translate into a distinct quality of intraoperative electrophysiological recordings, with an expected higher quality of STN signal detected in the central channel of the side with the better targeting accuracy (in this case, the first side). In this situation, another

expected observation is a different magnitude of mismatch between the initially planned versus final electrode positions, with a greater deviation on the second side.

Materials and methods

Study population

This observational retrospective cohort study included all consecutive patients who underwent simultaneous bilateral DBS surgery for the treatment of PD, with the STN as the target, from January 2019 to February 2021, and who had at least one month of follow-up. Patients who did not have a CT scan taken at least a month after the surgery, or who lacked data about the initial stereotactic planning, were excluded from this analysis. Patient data was collected from medical records in our database. The research protocol was approved by the ethics committee of the São João Hospital Centre.

Surgical procedure

Stereotactic planning was done by fusing a preoperative 1.5 T or 3 T MRI with a stereotactic CT scan obtained on the morning of the surgery, using FrameLink StealthStation 8[®] (Medtronic, USA), as previously described.²⁰ After verification of fusion accuracy, the mid AC-PC point was used to locate the anatomically defined target (12 mm lateral, 2 mm posterior and 4 mm inferior). The final position of the target was confirmed through direct visualization of the STN in T2-weighted or T2 SPACE images, followed by manual adjustment.

The surgery was performed with the patient supine, with a 30° head flexion, and under local anaesthesia. The first side operated was always contralateral to the most symptomatic side.

Whenever possible, three steel cannulas and microelectrodes (central, anterior and lateral) were inserted to perform stimulation. MER was started 5 mm above the planned target, advanced in 0.5 mm steps, and extended 3-5 mm below the target point or until the substantia nigra was detected, as previously described.²¹ Two experienced neurologists performed a visual and sound analysis of the single unit recordings captured by the high impedance electrodes, and qualitatively scored each 0.5 mm step using the following scale: 0 (no signal), 1 (low-quality: only traces visible), 2 (medium-quality: sparse activity) or 3 (high-quality: abundant STN-typical spikes). The STN is characterized by a mean firing rate of 37 ± 17 Hz, large amplitude of spikes and irregular rhythm.²² We used this information to estimate the length of high-quality STN activity detected along each channel. In case of poor MER results, lack of motor benefit, or presence of adverse effects in the original three trajectories (central, anterior and lateral), 2 additional cannulas in the medial and posterior channels were implanted in order to probe the quality of these regions. The channel and the final depth of implantation (defined as position of the lowermost contact) of each electrode were decided based on MER and macrostimulation test results. In our cohort, most patients received Activa[™] PC or Percept[™] PC (Medtronic, USA) IPGs, while a few received Vercise[™] PC (Boston Scientific, USA) IPGs.




Fig. 1 – Example MRI showing the trajectories of the implanted electrodes in the coronal (a) and axial (b) planes. The electrode artefacts visible in the post-operative CT scan were used to determine the entry point and the distal tip.

After the surgery, a CT scan was performed and fused with the preoperative MRI to evaluate the final position of the electrodes and exclude complications such as haemorrhage or ischaemia. In this study, this same analysis was repeated at least a month after the surgery, making use of the FrameLink StealthStation 8[®] software to determine the final coordinates of the tip of the electrode (shown in Fig. 1).

Statistical analysis

Statistical analysis and graphs were performed using GraphPad Prism (Version 9.0.0, GraphPad Software, San Diego, California, USA) and IBM SPSS Statistics for Windows (Version 27.0, IBM, Armonk, New York, USA). The Kolmogorov–Smirnov test was used to assess if the data had a normal distribution. McNemar’s test was used to analyze differences in the chosen electrode trajectory (central vs. non-central) between the first and second sides, and the left and right hemispheres. Using the initially planned coordinates from saved surgery plans, and the final electrode locations, the Euclidean vector deviation was calculated using Pythagoras’ theorem applied to three-dimensional space. Displacements in the x-, y- and z-directions were calculated by subtracting final and planned coordinates in each respective plane. Paired t-test was then used to evaluate differences between the first and second sides. Regarding the length of high-quality STN activity on each channel, comparisons were done using the Wilcoxon matched-pairs signed-rank test. Categorical variables were expressed as percentages and continuous variables as mean \pm standard deviation. The level of significance considered was $p < 0.05$.

Results

Study population

Of the 52 patients who received bilateral DBS surgery between January 2019 and February 2021, only 30 patients (60 implanted electrodes) fulfilled all the criteria and were included in this study.

Sixteen (53%) of the patients were female, the mean age at the time of surgery was 62.43 ± 6.63 years and the mean

Table 1 – Patient demographics.

Number of patients	30
Female patients, n (%)	16 (53%)
Age at the time of surgery, mean years \pm SD	62.4 ± 6.63
Duration of disease, mean years \pm SD	11.3 ± 5.74
Patients with right side as first implanted side, n (%)	16 (53%)
UPDRS-III med-off before surgery, mean \pm SD	54.6 ± 14.8
UPDRS-III med-off 1 month after surgery, mean \pm SD	13.5 ± 7.49

duration of disease was 11.3 ± 5.74 years. The mean UPDRS-III score off medication before surgery was 54.6 ± 14.8 . One month after surgery, the mean UPDRS-III score off medication and with stimulation switched on was 13.5 ± 7.49 . Sixteen patients (53%) had the first electrode implanted on the right side (shown in Table 1).

Electrode trajectories

Overall, the central channel was the most often used, in a total of 33 electrodes (55%), followed by the anterior channel, chosen for 21 electrodes (35%). The lateral channel was used in 4 cases and the medial in 2. The trajectory used for the implantation of the electrodes was symmetrical in 43% of patients.

As shown in Fig. 2a and b, the final leads were implanted in the central channel more frequently on the first side: in 21 patients (70.0%) for the first side and in 12 patients (40.0%) for the second ($p = 0.022$). Furthermore, regarding the first implanted side, the anterior channel was chosen in 7 patients (23.3%), while the lateral and medial channels were chosen in only 1 patient each. Concerning the second side, the anterior channel was used in 14 patients (46.7%), the lateral in 3 (10.0%) and the medial in 1 (3.33%). Cases where the medial channel was chosen were due to the low threshold for adverse effects in the other channels. The posterior channel was never used on either side.

When comparing the electrode trajectories chosen on the left and right cerebral hemispheres, regardless of which side was first implanted, the central channel was chosen in 60% and 50% of the cases, the anterior in 23.3% and 46.7%, and the lateral in 10.0% and 3.33%, respectively. The medial channel was only used in the left hemisphere. The differences between both hemispheres were not statistically significant ($p = 0.58$; shown in Fig. 2c and d).

STN activity detected by MER

Regarding the first side, MER activity was observed in at least one channel in all patients; however, there were 5 patients where not all channels were tested (1 anterior, 3 lateral, and 1 anterior and lateral). On the second side, all channels were tested in every patient, but there were 3 cases where there was no high-quality MER activity in any of them. In these situations, the final trajectories chosen were based on the macrostimulation test results alone.

The mean length of high-quality STN activity recorded in each channel is shown in Table 2. For the central channel, the first side showed a longer length of activity than the second, but the difference wasn’t significant (3.07 ± 1.85 mm




Fig. 2 – Frequency distribution of final electrode trajectories on the first side (a) and second side (b), and on the left hemisphere (c) and right hemisphere (d).

Table 2 – High-quality electrophysiological activity of the STN recorded during MER.

	First side	Second side	p-Value
Central channel, n	30	30	
Mean length, mm ± SD	3.07 ± 1.85	2.75 ± 1.94	0.539
No STN activity, n (%)	5 (17)	7 (23)	
Anterior channel, n	28	30	
Mean length, mm ± SD	1.59 ± 2.07	2.78 ± 2.14	0.0367
No STN activity, n (%)	14 (50)	9 (30)	
Lateral channel, n	26	30	
Mean length, mm ± SD	2.10 ± 1.94	2.15 ± 1.97	0.775
No STN activity, n (%)	10 (38)	11 (37)	
Extra channels			
Medial channel, n (%)	1 (3)	1 (3)	

vs. 2.75 ± 1.94 mm; $p=0.539$). On the other hand, the second side showed a significant increase in activity on the anterior channel when compared to the first (2.78 ± 2.14 mm vs. 1.59 ± 2.07 mm, respectively; $p=0.0367$). In the lateral channel, both sides had a similar length of STN activity (2.10 ± 1.94 mm

on the first side vs. 2.15 ± 1.97 mm on the second; $p=0.775$). However, the final trajectory chosen did not always coincide with the channel that showed the longest MER activity, as this decision was also influenced by the results of intraoperative macrostimulation. The final trajectory chosen corresponded




Fig. 3 – Mean Euclidean vector deviation \pm SD between planned target and final electrode position on the first and second sides (not including direction).

with the one with better MER results in 19 cases (63%) on the first side and 17 cases (57%) on the second side.

Planned versus final electrode position mismatch

As shown in Fig. 3, when comparing the 3D Euclidean vector deviation, regardless of the direction of movement, there was no statistically significant difference between both sides (2.45 ± 0.921 mm on the first side vs. 2.89 ± 1.37 mm on the second side; $p = 0.0881$). However, when comparing separate displacements in the x-, y- and z-directions, we found that electrodes implanted on the first side were placed on average 0.178 ± 0.917 mm lateral, 0.126 ± 1.10 mm posterior and 1.48 ± 1.64 mm inferior to the planned target, while the electrodes placed on the second side were 0.251 ± 1.08 mm medial, 0.355 ± 1.29 mm anterior and 2.26 ± 1.47 mm inferior to the initial target (shown in Fig. 4a and b). The differences in deviation in the y-direction and z-direction were statistically significant ($p = 0.0443$ and $p = 0.0029$, respectively), unlike those in the x-direction ($p = 0.104$).

Discussion

Electrode positioning

Our results showed a tendency to opt for the central channel more frequently on the first side than on the second. This difference was not attributable to a specific brain hemisphere, but rather to the order by which the electrodes were implanted. Two studies reported a similar trend, with the central channel being used more on the first side than the second, while the opposite occurred with the anterior channel.^{21,23} Others showed that, despite the central trajectory prevailing on both sides, the percentage of cases where the anterior trajectory was chosen was higher on the second side.^{24–26} Conversely, in a study by Umemura et al., 21% of the electrodes were placed in a trajectory other than the central. In these cases, the posterior channel was the most frequently chosen, followed by the anterior channel.²⁷ In another study, MER results led to a medial and posterior correction of 35.7% of the original stereotactic

coordinates obtained by CT-MRI image fusion.²⁸ In these studies, brain shift played a role in the differences observed.

STN activity detected by MER

In line with our initial hypothesis, the mean length of high-quality STN activity detected in our cohort was larger in the central channel on the first side, while in the anterior channel, better recordings were obtained on the second side. This supports the case for a lower accuracy in the final positioning of the second implanted electrode, in relation to the initially planned target.

Planned versus final electrode position mismatch

Our results showed no statistically significant differences regarding the Euclidean vector displacement, which is in agreement with results from other authors.^{6,10} Nevertheless, when analysing displacement in each axis, we found a tendency for the second electrode to be positioned more anterior and inferior to the planned target than the first one, further supporting our hypothesis. Other studies reported similar results.^{6,7,18} Slotty et al. also showed that the highest median deviation on both sides was seen in the z-axis as it is prone to larger changes depending on the results of MER.¹⁰ In our cohort, the larger deviation in the z-axis is probably due to the use of directional leads, which must be implanted at a greater depth than the one initially planned to ensure that contacts are placed on the target. Still, it should be emphasized that changes in the depth of the electrode are not equal to changes in the z-axis. Movement along the trajectory of the electrode will affect the target position in all axis, depending on the angle of insertion.¹⁷

Brain shift has been appointed as a cause for the differences observed between the first and second operated sides. According to this hypothesis, one would expect the STN on the second side to shift posteriorly and, therefore, lower quality MER activity on the anterior channel and, consequently, this trajectory to be chosen less frequently. However, our results show exactly the opposite pattern, which is in line with studies that have shown that brain shift in subcortical structures is very limited.^{10,15} Supporting the external validity of our data, two other studies have reported similar results, with an increased use of anterior trajectories on the second side.^{21,23} However, they offer no probable explanation for this difference. Chrastina et al. reported that, despite the central trajectory prevailing on both sides, the anterior trajectory was still chosen more frequently on the second side than on the first. They suggest that causes other than brain shift must play a role.²⁴

Consequently, other factors beyond pneumocephalus should be taken into consideration. A hypothesis is image distortion due to MRI, since it does not provide an accurate representation of the electrophysiological boundaries of the STN, therefore we should not rely solely on this information. For this reason, we fused MRI and CT images. However, since the pre-operative MRI is done with the patient supine, the true position of the STN during surgery might no longer be accurately represented by the planned target, as the patient is placed at a 30° angle. Additional sources of error include




Fig. 4 – Mean distance \pm SD between the planned target and the final electrode position in the x- and y-planes (a) and in the y- and z-planes (b), on the first and second sides.

imaging discrepancies, errors in target selection, and vector calculations.⁸ Surgical factors, such as dislocation or mechanical inaccuracy of the stereotactic frame, misinterpretation of the MER data, or factors related to the patient or the inherent disease, such as age or disease duration, can also play a role in these alterations.²⁴

Study limitations

The present study has some limitations. First, its retrospective nature and the small sample size mean that any extrapolations from our findings should be taken with caution, although the high consistency between our electrophysiological and anatomical approaches provides an important degree of robustness to our conclusions. Second, the identification of the final electrode position was done manually, by direct visualization in post-operative CT scans. This too can be a source of error, as the diameter of the leads is 1.27 mm, but the artefact can measure up to 3.3 mm, hindering the identification of the centre of the most distal tip of the lead.²⁹ Furthermore, the CT scan images used have a slice thickness of 1 mm so any values below this cannot be considered accurate. Third, the stereotactic techniques used can be responsible for a Euclidean distance of up to 2 mm between the planned target and the final lead location,⁸ though more recent methods managed to obtain deviations smaller than 1.5 mm.³⁰ Our results fit this range of values, despite being larger than those obtained by some other studies.^{6,31}

Future studies should be conducted to compare electrode displacement between the planned target and the electrode artefact in immediate and in delayed post-operative images, as well as to evaluate the clinical significance of the differences observed, by comparing the preoperative and postoperative UPDRS-III scores on each implanted side.

Conclusion

This study showed that in bilateral STN-DBS there is a tendency for the placement of the first implanted electrode to be more accurate and for the second electrode to have a

statistically significant deviation in the anterior and inferior directions. The initial target planning could possibly benefit from previously placing the second electrode more anteriorly, therefore reducing the number of MER trajectories tested and, hence, the duration of the surgery. Additionally, these results suggest that brain shift cannot be considered the main cause for the differences observed in our cohort. Further studies, with larger sample sizes and prospective designs, are required to analyze the role played by other factors. Nevertheless, our findings should be taken into account during planning for the second implanted side, in order to achieve greater accuracy in the placement of electrodes.

Funding

This research did not receive any specific grant from funding agencies in the public, commercial, or not-for-profit sectors.

Conflict of interest

Nothing to declare.

Acknowledgements

The authors wish to thank the Neurosurgery and Neurology teams of the São João Hospital Centre, in particular the members of the Movement Disorders and Functional Neurosurgery Unit.

REFERENCES

- Herrington TM, Cheng JJ, Eskandar EN. Mechanisms of deep brain stimulation. *J Neurophysiol.* 2016;115:19–38.
- Chiken S, Nambu A. Mechanism of deep brain stimulation: inhibition, excitation, or disruption? *Neuroscientist.* 2016;22:313–22.

3. Deuschl G, Schade-Brittinger C, Krack P, Volkmann J, Schäfer H, Bötzel K, et al. A randomized trial of deep-brain stimulation for Parkinson's disease. *N Engl J Med*. 2006;355:896–908.
4. Erola T, Heikkinen ER, Haapaniemi T, Tuominen J, Juolasmaa A, Myllylä VV. Efficacy of bilateral subthalamic nucleus (STN) stimulation in Parkinson's disease. *Acta Neurochir (Wien)*. 2006;148:389–94.
5. Sillay KA, Kumbier LM, Ross C, Brady M, Alexander A, Gupta A, et al. Perioperative brain shift and deep brain stimulating electrode deformation analysis: implications for rigid and non-rigid devices. *Ann Biomed Eng*. 2013;41:293–304.
6. Sadeghi Y, Pralong E, Knebel JF, Vingerhoets F, Pollo C, Levivier M, et al. Bilateral deep brain stimulation: the placement of the second electrode is not necessarily less accurate than that of the first one. *Stereotact Funct Neurosurg*. 2015;93:160–7.
7. Azmi H, Machado A, Deogaonkar M, Rezai A. Intracranial air correlates with preoperative cerebral atrophy and stereotactic error during bilateral STN DBS surgery for Parkinson's disease. *Stereotact Funct Neurosurg*. 2011;89:246–52.
8. van den Munckhof P, Bot M, Schuurman PR. Targeting of the subthalamic nucleus in patients with Parkinson's disease undergoing deep brain stimulation surgery. *Neurol Ther*. 2021;10:61–73.
9. Khan MF, Mewes K, Gross RE, Skrinjar O. Assessment of brain shift related to deep brain stimulation surgery. *Stereotact Funct Neurosurg*. 2008;86:44–53.
10. Slotty PJ, Kamp MA, Wille C, Kinfe TM, Steiger HJ, Vesper J. The impact of brain shift in deep brain stimulation surgery: observation and obviation. *Acta Neurochir (Wien)*. 2012;154:2063–8 [discussion 2068].
11. Elias WJ, Fu KM, Frysinger RC. Cortical and subcortical brain shift during stereotactic procedures. *J Neurosurg*. 2007;107:983–8.
12. Miyagi Y, Shima F, Sasaki T. Brain shift: an error factor during implantation of deep brain stimulation electrodes. *J Neurosurg*. 2007;107:989–97.
13. Halpern CH, Danish SF, Baltuch GH, Jaggi JL. Brain shift during deep brain stimulation surgery for Parkinson's disease. *Stereotact Funct Neurosurg*. 2008;86:37–43.
14. Matias CM, Frizon LA, Asfahan F, Uribe JD, Machado AG. Brain shift and pneumocephalus assessment during frame-based deep brain stimulation implantation with intraoperative magnetic resonance imaging. *Oper Neurosurg (Hagerstown)*. 2018;14:668–74.
15. Petersen EA, Holl EM, Martinez-Torres I, Foltynie T, Limousin P, Hariz MI, et al. Minimizing brain shift in stereotactic functional neurosurgery. *Neurosurgery*. 2010;67 (3 Suppl Operative): ons213-221; discussion ons221.
16. Hunsche S, Sauner D, Maarouf M, Poggenborg J, Lackner K, Sturm V, et al. Intraoperative X-ray detection and MRI-based quantification of brain shift effects subsequent to implantation of the first electrode in bilateral implantation of deep brain stimulation electrodes. *Stereotact Funct Neurosurg*. 2009;87:322–9.
17. Hamid NA, Mitchell RD, Mocroft P, Westby GW, Milner J, Pall H. Targeting the subthalamic nucleus for deep brain stimulation: technical approach and fusion of pre- and postoperative MR images to define accuracy of lead placement. *J Neurol Neurosurg Psychiatry*. 2005;76:409–14.
18. Sammartino F, Krishna V, King NK, Bruno V, Kalia S, Hodaie M, et al. Sequence of electrode implantation and outcome of deep brain stimulation for Parkinson's disease. *J Neurol Neurosurg Psychiatry*. 2016;87:859–63.
19. Tu PH, Liu ZH, Chen CC, Lin WY, Bowes AL, Lu CS, et al. Indirect targeting of subthalamic deep brain stimulation guided by stereotactic computed tomography and microelectrode recordings in patients with Parkinson's disease. *Front Hum Neurosci*. 2018;12:470.
20. Monteiro A, Andrade C, Rosas MJ, Linhares P, Massano J, Vaz R, et al. Deep brain stimulation of the subthalamic nucleus in advanced Parkinson's disease: five year follow-up at a Portuguese center. *Rev Neurol*. 2014;58:433–9.
21. Soares MI, Soares-Dos-Reis R, Rosas MJ, Monteiro P, Massano J. Intraoperative microelectrode recording in Parkinson's disease subthalamic deep brain stimulation: analysis of clinical utility. *J Clin Neurosci*. 2019;69:104–8.
22. Hutchison WD, Allan RJ, Opitz H, Levy R, Dostrovsky JO, Lang AE, et al. Neurophysiological identification of the subthalamic nucleus in surgery for Parkinson's disease. *Ann Neurol*. 1998;44:622–8.
23. Frequin HL, Bot M, Dilai J, Scholten MN, Postma M, Bour LJ, et al. Relative contribution of magnetic resonance imaging, microelectrode recordings, and awake test stimulation in final lead placement during deep brain stimulation surgery of the subthalamic nucleus in Parkinson's disease. *Stereotact Funct Neurosurg*. 2020;98:118–28.
24. Chrastina J, Novák Z, Baláž M, Říha I, Bočková M, Rektor I. The role of brain shift, patient age, and Parkinson's disease duration in the difference between anatomical and electrophysiological targets for subthalamic stimulation. *Br J Neurosurg*. 2013;27:676–82.
25. Bour LJ, Contarino MF, Foncke EM, de Bie RM, van den Munckhof P, Speelman JD, et al. Long-term experience with intraoperative microrecording during DBS neurosurgery in STN and GPi. *Acta Neurochir (Wien)*. 2010;152:2069–77.
26. Amirnovin R, Williams ZM, Cosgrove GR, Eskandar EN. Experience with microelectrode guided subthalamic nucleus deep brain stimulation. *Neurosurgery*. 2006;58 Suppl. ONS96-102; discussion ONS196-102.
27. Umemura A, Oka Y, Yamada K, Oyama G, Shimo Y, Hattori N. Validity of single tract microelectrode recording in subthalamic nucleus stimulation. *Neurol Med Chir (Tokyo)*. 2013;53:821–7.
28. Lanotte MM, Rizzone M, Bergamasco B, Faccani G, Melcarne A, Lopiano L. Deep brain stimulation of the subthalamic nucleus: anatomical, neurophysiological, and outcome correlations with the effects of stimulation. *J Neurol Neurosurg Psychiatry*. 2002;72:53–8.
29. Hemm S, Coste J, Gabrillargues J, Ouchchane L, Sarry L, Caire F, et al. Contact position analysis of deep brain stimulation electrodes on post-operative CT images. *Acta Neurochir (Wien)*. 2009;151:823–9 [discussion 829].
30. Li Z, Zhang JG, Ye Y, Li X. Review on factors affecting targeting accuracy of deep brain stimulation electrode implantation between 2001 and 2015. *Stereotact Funct Neurosurg*. 2016;94:351–62.
31. Krauss P, Oertel MF, Baumann-Vogel H, Imbach L, Baumann CR, Sarnthein J, et al. Intraoperative neurophysiologic assessment in deep brain stimulation surgery and its impact on lead placement. *J Neurol Surg A Cent Eur Neurosurg*. 2021;82:18–26.